

Products: Stabilizer Range of Motion Walker

Models: 311, 313, 315, 317, 318

- Patented walker design offers a superior rocker bottom in a low profile walker
- Continuous arching rocker design promotes a smooth energy conserving gait
- Adjustable ankle motion ranges from fixed to 45 degrees
- Fits either left or right foot

Order#	Size	Men's Shoe Size	Women's Shoe Size
311	X-Small	Up to 3 ^{1/2}	Up to 4 ^{1/2}
313	Small	4 - 7	5 - 8 ^{1/2}
315	Medium	7 ^{1/2} - 10	9 - 11
317	Large	10 ^{1/2} - 12	11 ^{1/2} - 13
318	X Large	12 ^{1/2} +	13 ^{1/2} +



SIZING: Measure according to shoe size.

HOW TO APPLY:

1. Unfasten straps, open liner, remove pads located inside boot and set aside.
2. While seated, place leg in brace and position heel at the back of the walker. Brace uprights should fall mid-line with the ankle.
3. Wrap liner around the leg and foot ensuring the large ankle pad is positioned over the top of the foot. Secure hook to the liner.
4. Position the uprights parallel to the leg per healthcare provider's recommendation.
5. Remove plastic sleeves from the uprights to expose hook.
6. Press hook on inside of the upright to the liner.
7. Beginning with the toe strap, thread straps through links and work up the leg by securing each strap.
8. Fasten closure straps until all are secured to ensure the liner and uprights don't migrate. NOTE: Pads may be inserted at either side of the ankle or at heel for additional comfort. NOTE: Liner and strap adjustment may be necessary as swelling patterns change throughout.

RANGE OF MOTION (ROM) JOINT APPLICATION INSTRUCTIONS

ROM joint can be fixed between 45° of dorsiflexion and 45° of plantarflexion by increments of 7.5°.

1. To obtain desired position, loosen hex screw.
2. Pull out pins, adjust dial to position, and place the pins in desired slots.
3. Tighten hex screw to secure position of the ROM dials.

CARE INSTRUCTIONS:

To clean, simply wipe with a damp cloth. DO NOT submerge the liner in water. Air dry away from heat. Do not machine wash or dry. Do not bleach. Spot clean the brace with a damp cloth if needed.

CAUTION:

Wear as recommended. Always consult a physician when encountering persistent pain, discomfort, inflammation, or swelling.

SUGG HCPCS CODE: L4386 and L4387